

Oz Material Testing Pty Ltd
Unit 10/24 Vale St
Malaga
WA 6090



NATA Accredited
Accreditation Number 2377
Site Number 2370 & 2554

Accredited for compliance with ISO/IEC 17025 – Testing
 NATA is a signatory to the ILAC Mutual Recognition
 Arrangement for the mutual recognition of the
 equivalence of testing, medical testing, calibration,
 inspection, proficiency testing scheme providers and
 reference materials producers reports and certificates.

Attention: **Sid Kamble**

Report **1311318-W**

Project name

Received Date **Jan 13, 2026**

Client Sample ID			5g Primer Liquid L26-Ja0014654 Not Provided ¹²	Lithium Based Curing Agent Liquid L26-Ja0014656 Not Provided ¹²
Sample Matrix				
Eurofins Sample No.				
Date Sampled				
Test/Reference	LOR	Unit		
Volatile Organics				
1.1-Dichloroethane	0.001	mg/L	M15 < 1	G01 < 0.01
1.1-Dichloroethene	0.001	mg/L	M15 < 1	G01 < 0.01
1.1.1-Trichloroethane	0.001	mg/L	M15 < 1	G01 < 0.01
1.1.1.2-Tetrachloroethane	0.001	mg/L	M15 < 1	G01 < 0.01
1.1.2-Trichloroethane	0.001	mg/L	M15 < 1	G01 < 0.01
1.1.2.2-Tetrachloroethane	0.001	mg/L	M15 < 1	G01 < 0.01
1.2-Dibromoethane	0.001	mg/L	M15 < 1	G01 < 0.01
1.2-Dichlorobenzene	0.001	mg/L	M15 < 1	< 0.01
1.2-Dichloroethane	0.001	mg/L	M15 < 1	G01 < 0.01
1.2-Dichloropropane	0.001	mg/L	M15 < 1	G01 < 0.01
1.2.3-Trichloropropane	0.001	mg/L	M15 < 1	G01 < 0.01
1.2.4-Trimethylbenzene	0.001	mg/L	M15 < 1	< 0.01
1.3-Dichlorobenzene	0.001	mg/L	M15 < 1	< 0.01
1.3-Dichloropropane	0.001	mg/L	M15 < 1	G01 < 0.01
1.3.5-Trimethylbenzene	0.001	mg/L	M15 < 1	< 0.01
1.4-Dichlorobenzene	0.001	mg/L	M15 < 1	< 0.01
2-Butanone (MEK)	0.005	mg/L	M15 < 5	G01 < 0.05
2-Propanone (Acetone)	0.005	mg/L	M15 < 5	G01 < 0.05
4-Chlorotoluene	0.001	mg/L	M15 < 1	G01 < 0.01
4-Methyl-2-pentanone (MIBK)	0.005	mg/L	M15 < 5	G01 < 0.05
Allyl chloride	0.001	mg/L	M15 < 1	G01 < 0.01
Benzene	0.001	mg/L	M15 < 1	G01 < 0.01
Bromobenzene	0.001	mg/L	M15 < 1	G01 < 0.01
Bromochloromethane	0.001	mg/L	M15 < 1	G01 < 0.01
Bromodichloromethane	0.001	mg/L	M15 < 1	G01 < 0.01
Bromoform	0.001	mg/L	M15 < 1	G01 < 0.01
Bromomethane	0.005	mg/L	M15 < 5	G01 < 0.05
Carbon disulfide	0.001	mg/L	M15 < 1	G01 < 0.01
Carbon Tetrachloride	0.001	mg/L	M15 < 1	G01 < 0.01
Chlorobenzene	0.001	mg/L	M15 < 1	G01 < 0.01
Chloroethane	0.005	mg/L	M15 < 5	G01 < 0.05
Chloroform	0.005	mg/L	M15 < 5	G01 < 0.05
Chloromethane	0.005	mg/L	M15 < 5	G01 < 0.05
cis-1.2-Dichloroethene	0.001	mg/L	M15 < 1	G01 < 0.01
cis-1.3-Dichloropropene	0.001	mg/L	M15 < 1	G01 < 0.01
Dibromochloromethane	0.001	mg/L	M15 < 1	G01 < 0.01

Client Sample ID			5g Primer Liquid	Lithium Based Curing Agent Liquid
Sample Matrix			L26-Ja0014654	L26-Ja0014656
Eurofins Sample No.			Not Provided ¹²	Not Provided ¹²
Date Sampled				
Test/Reference	LOR	Unit		
Volatile Organics				
Dibromomethane	0.001	mg/L	M15 < 1	G01 < 0.01
Dichlorodifluoromethane	0.005	mg/L	M15 < 5	G01 < 0.05
Ethylbenzene	0.001	mg/L	M15 < 1	G01 < 0.01
Iodomethane	0.001	mg/L	M15 < 1	G01 < 0.01
Isopropyl benzene (Cumene)	0.001	mg/L	M15 < 1	G01 < 0.01
m&p-Xylenes	0.002	mg/L	M15 < 2	G01 < 0.02
Methylene Chloride	0.005	mg/L	M15 < 5	G01 < 0.05
o-Xylene	0.001	mg/L	M15 < 1	G01 < 0.01
Styrene	0.001	mg/L	M15 2.5	G01 < 0.01
Tetrachloroethene	0.001	mg/L	M15 < 1	G01 < 0.01
Toluene	0.001	mg/L	M15 < 1	G01 < 0.01
trans-1.2-Dichloroethene	0.001	mg/L	M15 < 1	G01 < 0.01
trans-1.3-Dichloropropene	0.001	mg/L	M15 < 1	G01 < 0.01
Trichloroethene	0.001	mg/L	M15 < 1	G01 < 0.01
Trichlorofluoromethane	0.005	mg/L	M15 < 5	G01 < 0.05
Vinyl chloride	0.005	mg/L	M15 < 5	G01 < 0.05
Xylenes - Total*	0.003	mg/L	M15 < 3	G01 < 0.03
Total MAH*	0.003	mg/L	2.5	< 0.02
4-Bromofluorobenzene (surr.)	1	%	115	110
Toluene-d8 (surr.)	1	%	108	110

Sample History

Where samples are submitted/analysed over several days, the last date of extraction is reported.

If the date and time of sampling are not provided, the Laboratory will not be responsible for compromised results should testing be performed outside the recommended holding time.

Description

Volatile Organics

- Method: ARL 132 - Purgeable Organics in Water by GCMS

Testing Site

Welshpool

Extracted

Jan 13, 2026

Holding Time

7 Days

Perth
 46-48 Banksia Road
 Welshpool
 WA 6106
 +61 8 6253 4444
 NATA# 2377
 Site# 2370 & 2554

Melbourne
 6 Monterey Road
 Dandenong South
 VIC 3175
 +61 3 8564 5000
 NATA# 1261
 Site# 1254

Geelong
 19/8 Lewalan Street
 Grovedale
 VIC 3216
 +61 3 8564 5000
 NATA# 1261
 Site# 25403

Sydney
 179 Magowar Road
 Gurraveen
 NSW 2145
 +61 2 9900 8400
 NATA# 1261
 Site# 18217

Canberra
 Unit 1,2 Dacre Street
 Mitchell
 ACT 2911
 +61 2 6113 8091
 NATA# 1261
 Site# 25466

Brisbane
 1/21 Smallwood Place
 Murarrie
 QLD 4172
 +61 7 3902 4600
 NATA# 1261
 Site# 20794 & 2780

Newcastle
 1/2 Frost Drive
 Mayfield West
 NSW 2304
 +61 2 4968 8448
 NATA# 1261
 Site# 25079

Company Name: Oz Material Testing Pty Ltd
Address: Unit 10/24 Vale St
 Malaga
 WA 6090

Order No.:
Report #: 1311318
Phone: 0416 394 904
Fax:
Received: Jan 13, 2026 2:55 PM
Due: Jan 20, 2026
Priority: 5 Day
Contact Name: Sid Kamble

Project Name:
Eurofins Analytical Services Manager : Douglas Todd

Sample Detail						Volatle Organics	Moisture Set
Perth Laboratory - NATA # 2377 Site # 2370 & 2554						X	X
External Laboratory							
No	Sample ID	Sample Date	Sampling Time	Matrix	LAB ID		
1	Feather Coad	Not Provided		Solid	L26-Ja0014651	X	X
2	Self Leveling Solid Compound	Not Provided		Solid	L26-Ja0014652	X	X
3	Topping	Not Provided		Solid	L26-Ja0014653	X	X
4	5g Primer	Not Provided		Liquid	L26-Ja0014654	X	
5	K10 Moisture Barrier	Not Provided		Solid	L26-Ja0014655	X	X
6	Lithium Based Curing Agent	Not Provided		Liquid	L26-Ja0014656	X	
Test Counts						6	4

Internal Quality Control Review and Glossary

General

- Laboratory QC results for Method Blanks, Duplicates, Matrix Spikes, and Laboratory Control Samples follow guidelines delineated in the National Environment Protection (Assessment of Site Contamination) Measure 1999, as amended May 2013. They are included in this QC report where applicable. Additional QC data may be available on request.
- Unless otherwise stated, all soil/sediment/solid results are reported on a dry weight basis.
- Unless otherwise stated, all biota/food results are reported on a wet weight basis on the edible portion.
- For CEC results where the sample's origin is unknown or environmentally contaminated, the results should be used advisedly.
- Actual LORs are matrix dependent. Quoted LORs may be raised where sample extracts are diluted due to interferences.
- Results are uncorrected for matrix spikes or surrogate recoveries except for PFAS compounds where annotated.
- SVOC analysis on waters is performed on homogenised, unfiltered samples unless noted otherwise.
- Samples were analysed on an 'as received' basis.
- Information identified in this report with **blue** colour indicates data provided by customers that may have an impact on the results.
- This report replaces any interim results previously issued.

Holding Times

Please refer to the 'Sample Preservation and Container Guide' for holding times (QS3001).

For samples received on the last day of holding time, notification of testing requirements should have been received at least 6 hours before sample receipt deadlines as stated on the SRA.

If the Laboratory did not receive the information in the required timeframe, and despite any other integrity issues, suitably qualified results may still be reported.

Holding times apply from the sampling date; therefore, compliance with these may be outside the laboratory's control.

For VOCs containing vinyl chloride, styrene and 2-chloroethyl vinyl ether, the holding time is seven days; however, for all other VOCs, such as BTEX or C6-10 TRH, the holding time is 14 days.

Units

mg/kg: milligrams per kilogram	mg/L: milligrams per litre	ppm: parts per million
µg/L: micrograms per litre	ppb: parts per billion	%: Percentage
org/100 mL: Organisms per 100 millilitres	NTU: Nephelometric Turbidity Units	MPN/100 mL: Most Probable Number of organisms per 100 millilitres
CFU: Colony Forming Unit	Colour: Pt-Co Units (CU)	

Terms

APHA	American Public Health Association
CEC	Cation Exchange Capacity
COC	Chain of Custody
CP	Client Parent - QC was performed on samples pertaining to this report
CRM	Certified Reference Material (ISO17034) - reported as percent recovery.
Dry	Where moisture has been determined on a solid sample, the result is expressed on a dry weight basis.
Duplicate	A second piece of analysis from the same sample and reported in the same units as the result to show comparison.
LOR	Limit of Reporting.
LCS	Laboratory Control Sample - reported as percent recovery.
Method Blank	In the case of solid samples, these are performed on laboratory-certified clean sands and in the case of water samples, these are performed on de-ionised water.
NCP	Non-Client Parent - QC performed on samples not pertaining to this report, QC represents the sequence or batch that client samples were analysed within.
RPD	Relative Percent Difference between two Duplicate pieces of analysis.
SPIKE	Addition of the analyte to the sample and reported as percentage recovery.
SRA	Sample Receipt Advice
Surr - Surrogate	The addition of a similar compound to the analyte target is reported as percentage recovery. See below for acceptance criteria.
TBTO	Tributyltin oxide (<i>bis</i> -tributyltin oxide) - individual tributyltin compounds cannot be identified separately in the environment; however, free tributyltin was measured, and its values were converted stoichiometrically into tributyltin oxide for comparison with regulatory limits.
TCLP	Toxicity Characteristic Leaching Procedure
TEQ	Toxic Equivalency Quotient or Total Equivalence
QSM	US Department of Defense Quality Systems Manual Version 6.0
US EPA	United States Environmental Protection Agency
WA DWER	Sum of PFBA, PFPeA, PFHxA, PFHpA, PFOA, PFBS, PFHxS, PFOS, 6:2 FTSA, 8:2 FTSA

QC - Acceptance Criteria

The acceptance criteria should only be used as a guide and may be different when site-specific Sampling Analysis and Quality Plan (SAQP) have been implemented.

RPD Duplicates: Global RPD Duplicates Acceptance Criteria is ≤30%; however, the following acceptance guidelines are equally applicable:

Results <10 times the LOR:	No Limit
Results between 10-20 times the LOR:	RPD must lie between 0-50%
Results >20 times the LOR:	RPD must lie between 0-30%

NOTE: pH duplicates are reported as a range, not as RPD

Surrogate Recoveries: Recoveries must lie between 20-130% for Speciated Phenols & 50-150% for PFAS. SVOCs recoveries 20 – 150%, VOC recoveries 50 – 150%

PFAS field samples containing surrogate recoveries above the QC limit designated in QSM 6.0, where no positive PFAS results have been reported or reviewed, and no data was affected.

QC Data General Comments

- Where a result is reported as less than (<), higher than the nominated LOR, this is due to either matrix interference, extract dilution required due to interferences or contaminant levels within the sample, high moisture content or insufficient sample provided.
- Duplicate data shown within this report that states the word "BATCH" is a Batch Duplicate from outside of your sample batch but within the laboratory sample batch at a 1:10 ratio. The Parent and Duplicate data shown are not data from your samples.
- pH and Free Chlorine analysed in the laboratory - Analysis on this test must begin within 30 minutes of sampling. Therefore, laboratory analysis is unlikely to be completed within holding time. Analysis will begin as soon as possible after sample receipt.
- Recovery Data (Spikes & Surrogates) - where chromatographic interference does not allow the determination of recovery, the term "INT" appears against that analyte.
- For Matrix Spikes and LCS results, a dash "-" in the report means that the specific analyte was not added to the QC sample.
- Duplicate RPDs are calculated from raw analytical data; thus, it is possible to have two sets of data.

Quality Control Results

Test	Units	Result 1			Acceptance Limits	Pass Limits	Qualifying Code
Method Blank							
Volatile Organics							
1.1-Dichloroethane	mg/L	< 0.001			0.001	Pass	
1.1-Dichloroethene	mg/L	< 0.001			0.001	Pass	
1.1.1-Trichloroethane	mg/L	< 0.001			0.001	Pass	
1.1.1.2-Tetrachloroethane	mg/L	< 0.001			0.001	Pass	
1.1.2-Trichloroethane	mg/L	< 0.001			0.001	Pass	
1.1.2.2-Tetrachloroethane	mg/L	< 0.001			0.001	Pass	
1.2-Dibromoethane	mg/L	< 0.001			0.001	Pass	
1.2-Dichlorobenzene	mg/L	< 0.001			0.001	Pass	
1.2-Dichloroethane	mg/L	< 0.001			0.001	Pass	
1.2-Dichloropropane	mg/L	< 0.001			0.001	Pass	
1.2.3-Trichloropropane	mg/L	< 0.001			0.001	Pass	
1.2.4-Trimethylbenzene	mg/L	< 0.001			0.001	Pass	
1.3-Dichlorobenzene	mg/L	< 0.001			0.001	Pass	
1.3-Dichloropropane	mg/L	< 0.001			0.001	Pass	
1.3.5-Trimethylbenzene	mg/L	< 0.001			0.001	Pass	
1.4-Dichlorobenzene	mg/L	< 0.001			0.001	Pass	
2-Butanone (MEK)	mg/L	< 0.005			0.005	Pass	
2-Propanone (Acetone)	mg/L	< 0.005			0.005	Pass	
4-Chlorotoluene	mg/L	< 0.001			0.001	Pass	
4-Methyl-2-pentanone (MIBK)	mg/L	< 0.005			0.005	Pass	
Allyl chloride	mg/L	< 0.001			0.001	Pass	
Benzene	mg/L	< 0.001			0.001	Pass	
Bromobenzene	mg/L	< 0.001			0.001	Pass	
Bromochloromethane	mg/L	< 0.001			0.001	Pass	
Bromodichloromethane	mg/L	< 0.001			0.001	Pass	
Bromoform	mg/L	< 0.001			0.001	Pass	
Bromomethane	mg/L	< 0.005			0.005	Pass	
Carbon disulfide	mg/L	< 0.001			0.001	Pass	
Carbon Tetrachloride	mg/L	< 0.001			0.001	Pass	
Chlorobenzene	mg/L	< 0.001			0.001	Pass	
Chloroethane	mg/L	< 0.005			0.005	Pass	
Chloroform	mg/L	< 0.005			0.005	Pass	
Chloromethane	mg/L	< 0.005			0.005	Pass	
cis-1.2-Dichloroethene	mg/L	< 0.001			0.001	Pass	
cis-1.3-Dichloropropene	mg/L	< 0.001			0.001	Pass	
Dibromochloromethane	mg/L	< 0.001			0.001	Pass	
Dibromomethane	mg/L	< 0.001			0.001	Pass	
Dichlorodifluoromethane	mg/L	< 0.005			0.005	Pass	
Ethylbenzene	mg/L	< 0.001			0.001	Pass	
Iodomethane	mg/L	< 0.001			0.001	Pass	
Isopropyl benzene (Cumene)	mg/L	< 0.001			0.001	Pass	
m&p-Xylenes	mg/L	< 0.002			0.002	Pass	
Methylene Chloride	mg/L	< 0.005			0.005	Pass	
o-Xylene	mg/L	< 0.001			0.001	Pass	
Styrene	mg/L	< 0.001			0.001	Pass	
Tetrachloroethene	mg/L	< 0.001			0.001	Pass	
Toluene	mg/L	< 0.001			0.001	Pass	
trans-1.2-Dichloroethene	mg/L	< 0.001			0.001	Pass	
trans-1.3-Dichloropropene	mg/L	< 0.001			0.001	Pass	
Trichloroethene	mg/L	< 0.001			0.001	Pass	

Test	Units	Result 1			Acceptance Limits	Pass Limits	Qualifying Code
Trichlorofluoromethane	mg/L	< 0.005			0.005	Pass	
Vinyl chloride	mg/L	< 0.005			0.005	Pass	
Xylenes - Total*	mg/L	< 0.003			0.003	Pass	
Method Blank							
Volatile Organics							
1.1-Dichloroethane	mg/L	< 0.001			0.001	Pass	
1.1-Dichloroethene	mg/L	< 0.001			0.001	Pass	
1.1.1-Trichloroethane	mg/L	< 0.001			0.001	Pass	
1.1.1.2-Tetrachloroethane	mg/L	< 0.001			0.001	Pass	
1.1.2-Trichloroethane	mg/L	< 0.001			0.001	Pass	
1.1.2.2-Tetrachloroethane	mg/L	< 0.001			0.001	Pass	
1.2-Dibromoethane	mg/L	< 0.001			0.001	Pass	
1.2-Dichlorobenzene	mg/L	< 0.001			0.001	Pass	
1.2-Dichloroethane	mg/L	< 0.001			0.001	Pass	
1.2-Dichloropropane	mg/L	< 0.001			0.001	Pass	
1.2.3-Trichloropropane	mg/L	< 0.001			0.001	Pass	
1.2.4-Trimethylbenzene	mg/L	< 0.001			0.001	Pass	
1.3-Dichlorobenzene	mg/L	< 0.001			0.001	Pass	
1.3-Dichloropropane	mg/L	< 0.001			0.001	Pass	
1.3.5-Trimethylbenzene	mg/L	< 0.001			0.001	Pass	
1.4-Dichlorobenzene	mg/L	< 0.001			0.001	Pass	
2-Butanone (MEK)	mg/L	< 0.005			0.005	Pass	
2-Propanone (Acetone)	mg/L	< 0.005			0.005	Pass	
4-Chlorotoluene	mg/L	< 0.001			0.001	Pass	
4-Methyl-2-pentanone (MIBK)	mg/L	< 0.005			0.005	Pass	
Allyl chloride	mg/L	< 0.001			0.001	Pass	
Benzene	mg/L	< 0.001			0.001	Pass	
Bromobenzene	mg/L	< 0.001			0.001	Pass	
Bromochloromethane	mg/L	< 0.001			0.001	Pass	
Bromodichloromethane	mg/L	< 0.001			0.001	Pass	
Bromoform	mg/L	< 0.001			0.001	Pass	
Bromomethane	mg/L	< 0.005			0.005	Pass	
Carbon disulfide	mg/L	< 0.001			0.001	Pass	
Carbon Tetrachloride	mg/L	< 0.001			0.001	Pass	
Chlorobenzene	mg/L	< 0.001			0.001	Pass	
Chloroethane	mg/L	< 0.005			0.005	Pass	
Chloroform	mg/L	< 0.005			0.005	Pass	
Chloromethane	mg/L	< 0.005			0.005	Pass	
cis-1.2-Dichloroethene	mg/L	< 0.001			0.001	Pass	
cis-1.3-Dichloropropene	mg/L	< 0.001			0.001	Pass	
Dibromochloromethane	mg/L	< 0.001			0.001	Pass	
Dibromomethane	mg/L	< 0.001			0.001	Pass	
Dichlorodifluoromethane	mg/L	< 0.005			0.005	Pass	
Ethylbenzene	mg/L	< 0.001			0.001	Pass	
Iodomethane	mg/L	< 0.001			0.001	Pass	
Isopropyl benzene (Cumene)	mg/L	< 0.001			0.001	Pass	
m&p-Xylenes	mg/L	< 0.002			0.002	Pass	
Methylene Chloride	mg/L	< 0.005			0.005	Pass	
o-Xylene	mg/L	< 0.001			0.001	Pass	
Styrene	mg/L	< 0.001			0.001	Pass	
Tetrachloroethene	mg/L	< 0.001			0.001	Pass	
Toluene	mg/L	< 0.001			0.001	Pass	
trans-1.2-Dichloroethene	mg/L	< 0.001			0.001	Pass	
trans-1.3-Dichloropropene	mg/L	< 0.001			0.001	Pass	

Test				Units	Result 1			Acceptance Limits	Pass Limits	Qualifying Code	
Trichloroethene				mg/L	< 0.001			0.001	Pass		
Trichlorofluoromethane				mg/L	< 0.005			0.005	Pass		
Vinyl chloride				mg/L	< 0.005			0.005	Pass		
Xylenes - Total*				mg/L	< 0.003			0.003	Pass		
LCS - % Recovery											
Volatile Organics											
1.1-Dichloroethene				%	91			70-130	Pass		
1.2-Dichlorobenzene				%	100			70-130	Pass		
Benzene				%	94			70-130	Pass		
Ethylbenzene				%	101			70-130	Pass		
m&p-Xylenes				%	105			70-130	Pass		
o-Xylene				%	104			70-130	Pass		
Toluene				%	101			70-130	Pass		
Trichloroethene				%	120			70-130	Pass		
Xylenes - Total*				%	105			70-130	Pass		
LCS - % Recovery											
Volatile Organics											
1.2-Dichloroethane				%	95			70-130	Pass		
Ethylbenzene				%	79			70-130	Pass		
m&p-Xylenes				%	81			70-130	Pass		
o-Xylene				%	80			70-130	Pass		
Toluene				%	79			70-130	Pass		
Trichloroethene				%	86			70-130	Pass		
Xylenes - Total*				%	81			70-130	Pass		
Test	Lab Sample ID	QA Source	Units	Result 1				Acceptance Limits	Pass Limits	Qualifying Code	
Spike - % Recovery											
Volatile Organics											
1.2-Dichloroethane				L26-Ja0025433	NCP	%	95		70-130	Pass	
Benzene				L26-Ja0022335	NCP	%	100		70-130	Pass	
Ethylbenzene				L26-Ja0022335	NCP	%	107		70-130	Pass	
m&p-Xylenes				L26-Ja0022335	NCP	%	110		70-130	Pass	
o-Xylene				L26-Ja0022335	NCP	%	112		70-130	Pass	
Toluene				L26-Ja0022335	NCP	%	105		70-130	Pass	
Trichloroethene				L26-Ja0025433	NCP	%	76		70-130	Pass	
Xylenes - Total*				L26-Ja0022335	NCP	%	110		70-130	Pass	
Test	Lab Sample ID	QA Source	Units	Result 1				Acceptance Limits	Pass Limits	Qualifying Code	
Duplicate											
Volatile Organics											
1.1-Dichloroethane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.1-Dichloroethene				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.1.1-Trichloroethane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.1.1.2-Tetrachloroethane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.1.2-Trichloroethane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.1.2.2-Tetrachloroethane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.2-Dibromoethane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.2-Dichlorobenzene				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.2-Dichloroethane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.2-Dichloropropane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.2.3-Trichloropropane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.2.4-Trimethylbenzene				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.3-Dichlorobenzene				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.3-Dichloropropane				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass
1.3.5-Trimethylbenzene				L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass

Test	Lab Sample ID	QA Source	Units	Result 1			Acceptance Limits	Pass Limits	Qualifying Code
Duplicate									
Volatile Organics				Result 1	Result 2	RPD			
1,4-Dichlorobenzene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
2-Butanone (MEK)	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
2-Propanone (Acetone)	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
4-Chlorotoluene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
4-Methyl-2-pentanone (MIBK)	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
Allyl chloride	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Benzene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Bromobenzene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Bromochloromethane	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Bromodichloromethane	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Bromoform	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Bromomethane	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
Carbon disulfide	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Carbon Tetrachloride	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Chlorobenzene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Chloroethane	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
Chloroform	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
Chloromethane	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
cis-1,2-Dichloroethene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
cis-1,3-Dichloropropene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Dibromochloromethane	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Dibromomethane	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Dichlorodifluoromethane	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
Ethylbenzene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Iodomethane	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Isopropyl benzene (Cumene)	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
m&p-Xylenes	L26-Ja0014656	CP	mg/L	< 0.02	< 0.002	<1	30%	Pass	
Methylene Chloride	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
o-Xylene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Styrene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Tetrachloroethene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Toluene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
trans-1,2-Dichloroethene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
trans-1,3-Dichloropropene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Trichloroethene	L26-Ja0014656	CP	mg/L	< 0.01	< 0.001	<1	30%	Pass	
Trichlorofluoromethane	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
Vinyl chloride	L26-Ja0014656	CP	mg/L	< 0.05	< 0.005	<1	30%	Pass	
Xylenes - Total*	L26-Ja0014656	CP	mg/L	< 0.03	< 0.003	<1	30%	Pass	

Comments
Sample Integrity

Custody Seals Intact (if used)	N/A
Attempt to Chill was evident	No
Sample correctly preserved	Yes
Appropriate sample containers have been used	Yes
Sample containers for volatile analysis received with minimal headspace	No
Samples received within HoldingTime	N/A
Some samples have been subcontracted	No

Qualifier Codes/Comments

Code	Description
G01	The LORs have been raised due to matrix interference
I12	Where sampling date has not been provided, Eurofins Environment Testing is not able to determine whether analysis has been performed within recommended holding times.
M15	LOR raised due to physical properties of sample

Authorised by:

Douglas Todd	Analytical Services Manager
Douglas Todd	Senior Analyst-Volatile



Kim Rodgers
General Manager

Final Report – this report replaces any previously issued Report

- Indicates Not Requested

* Indicates NATA accreditation does not cover the performance of this service

Measurement uncertainty of test data is available on request

Eurofins shall not be liable for loss, cost, damages or expenses incurred by the client, or any other person or company, resulting from the use of any information or interpretation given in this report. In no case shall Eurofins be liable for consequential damages including, but not limited to, lost profits, damages for failure to meet deadlines and lost production arising from this report. This document shall not be reproduced except in full and relates only to the items tested. Unless indicated otherwise, the tests were performed on the samples as received.